



Sterile Single Use Forceps Instructions for Use

Sterile Single-Use Manual Forceps Intended Use

Sterile single-use manual forceps are intended for use in multiple surgical specialties to perform basic surgical tasks such as grasping or clamping.

Indications For Use

These forceps are intended to grasp, hold or manipulate tissue or structures. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The forceps consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use forceps in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Tweezers

Instructions for Use

Sterile Single-Use Manual Tweezers Intended Use

Sterile single-use manual tweezers are intended for use in multiple surgical specialties to perform basic surgical tasks such as grasping or clamping.

Indications For Use

These tweezers are intended to grasp, hold or manipulate tissue or structures. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The tweezers consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use tweezers in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Scissors Instructions for Use

Sterile Single-Use Manual Scissors Intended Use

Sterile single-use manual scissors are intended for use in multiple surgical specialties to perform basic surgical tasks such as cutting.

Indications For Use

These Scissors are intended to be used for cutting sutures, dressings, and cutting and dissecting biological tissue. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The Scissors consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Scissors in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Phoenix Surgical Supplies must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Phoenix Surgical Supplies within thirty (30) days of Phoenix Surgical Supplies receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Needle Holder Instructions for Use

Sterile Single-Use Manual Needle Holders Intended Use

Sterile single-use manual Needle Holders are intended for use in multiple surgical specialties to perform basic surgical tasks such as manipulating and grasping needles.

Indications For Use

A needle holder, also called needle driver or needle forceps, can be used to hold and push a suturing needle when performing wound closure, ligation and other surgical procedures that require re-anastomosis. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The Needle Holders consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Needle Holders in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Scalpel Handle Instructions for Use

Sterile Single-Use Manual Retractor Intended Use

Sterile single-use manual Scalpel Handle are intended to be used to hold scalpel blades securely throughout any procedure a healthcare professional may be carrying out.

Indications For Use

This device is intended to be used to hold scalpel blades of varying sizes and designs. The blade is locked in place by using a key-slot type mechanism which safely holds the blade in place. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Scalpel Handles consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Biopsy Punch Instructions for Use

Sterile Single-Use Manual Biopsy Punches Intended Use

Sterile single-use manual Biopsy Punches are intended for use in multiple surgical specialties to perform basic surgical tasks such as the removal of a small regular sample of tissue.

Indications For Use

These Biopsy Punches are intended to obtain tissue samples for diagnosis and treatment in a variety of procedures, including skin biopsies and cervical biopsies. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The Biopsy Punches consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Biopsy Punches in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Retractor Instructions for Use

Sterile Single-Use Manual Retractor Intended Use

Sterile single-use manual Retractors are intended to be used to separate the edges of a surgical incision/wound or to hold away certain organs and tissues

Indications For Use

This device is intended to be used to retract skin, fat, and muscle during procedures such as hernia, breast, axillary surgery, or orthopaedic operations. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Retractors consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Staple Remover Instructions for Use

Sterile Single-Use Manual Staple Remover Intended Use

Sterile single-use manual Staple Removers are intended to be used to provide fast and easy removal of all types of surgical skin staples.

Indications For Use

This device is intended to be used to provide fast and easy removal of all types of surgical skin staples. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Staple Removers consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Wire Cutter Instructions for Use

Sterile Single-Use Manual Wire Cutter Intended Use

Sterile single-use manual Wire Cutters are intended to be used to cut wires and pins used in procedures such as orthopaedic surgeries.

Indications For Use

This device is intended to be used to cut wires and pins used in procedures such as orthopaedic surgeries. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Wire Cutters consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Nasal Speculum Instructions for Use

Sterile Single-Use Manual Nasal Speculum Intended Use

Sterile single-use manual Nasal Speculums are intended for use in a range of procedures, such as anterior rhinoscopy, septal surgeries and the removal of foreign bodies and polyps.

Indications For Use

Nasal Speculums are indicated for use in diagnosis, extraction of foreign bodies, and nasal packing by healthcare professionals on all patient populations. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Nasal Speculum consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Vaginal Speculum Instructions for Use

Sterile Single-Use Manual Vaginal Speculum Intended Use

Sterile single-use manual Vaginal Speculums are intended for opening the vagina to expose the interior of the vagina

Indications For Use

Vaginal Speculums are indicated for use in diagnosis, extraction of foreign bodies, and Vaginal packing by healthcare professionals on all patient populations. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Vaginal Speculums consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Rectal Speculum Instructions for Use

Sterile Single-Use Manual Rectal Speculum Intended Use

Sterile single-use manual Rectal Speculums are intended for use in dilating the anus and allowing examination of a patient's anal canal.

Indications For Use

Rectal Speculums are indicated for use in dilating the anus and allowing examination of a patient's anal canal. It is the responsibility of the surgical team to select the appropriate instrument for each case.

Materials

The Rectal Speculums consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Sterile Single Use Proctoscope Instructions for Use

Sterile Single-Use Manual Proctoscope Intended Use

Sterile single-use manual Proctoscope are intended to be inserted into and to open the anus.

Indications For Use

This device is intended to be used for some necessary examinations used in the field of Proctology and/or Urology. It is the responsibility of healthcare professionals to select the appropriate instrument for each case.

Materials

The Proctoscopes consist of the following patient contacting materials: stainless steel

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use Rectal Speculum in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

Fremantle Stainless disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to Fremantle Stainless. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a Fremantle Stainless Product (hereinafter the "Product") that should the Product fail to function to Fremantle Stainless' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), Fremantle Stainless will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from Fremantle Stainless or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). Fremantle Stainless must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to Fremantle Stainless within thirty (30) days of Fremantle Stainless receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, Fremantle Stainless shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than Fremantle Stainless or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall Fremantle Stainless be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. Fremantle Stainless reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Single Use Sigmoidoscope Instructions for Use

Sterile Single-Use Sigmoidoscopes Intended Use

This device is a special tube used by inserting into and through the anus.

Indications For Use

This device is intended to be used for examining the interior of the rectum and sigmoid /lower colon.

Materials

The sigmoidoscopes consist of the following patient contacting materials: High-density polyethylene

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use sigmoidoscopes in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

North EOS Industries Inc. disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to North EOS Industries Inc. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a North EOS Industries Inc Product (hereinafter the "Product") that should the Product fail to function to North EOS Industries Inc' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), North EOS Industries Inc will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from North EOS Industries Inc or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). North EOS Industries Inc must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to North EOS Industries Inc within thirty (30) days of North EOS Industries Inc receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, North EOS Industries Inc shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than North EOS Industries Inc or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall North EOS Industries Inc be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. North EOS Industries Inc reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Single Use Proctoscopes Instructions for Use

Sterile Single-Use Proctoscopes Intended Use

This device is to insert into and open the anus.

Indications For Use

This device is intended to be used for some necessary examinations used in the field of Proctology and/or Urology.

Materials

The proctoscopes consist of the following patient contacting materials: High-density polyethylene

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use proctoscopes in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

North EOS Industries Inc. disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to North EOS Industries Inc. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a North EOS Industries Inc Product (hereinafter the "Product") that should the Product fail to function to North EOS Industries Inc' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), North EOS Industries Inc will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from North EOS Industries Inc or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). North EOS Industries Inc must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to North EOS Industries Inc within thirty (30) days of North EOS Industries Inc receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, North EOS Industries Inc shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than North EOS Industries Inc or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall North EOS Industries Inc be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. North EOS Industries Inc reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS



Single Use Vaginal Speculums Instructions for Use

Sterile Single-Use Vaginal Speculums Intended Use

This device is for opening the vagina to expose the interior of the vagina..

Indications For Use

This device is intended to be used to perform routine examinations or take cytological samples from the surface of the vagina or cervix.

Materials

The vaginal speculums consist of the following patient contacting materials: High-density polyethylene

Contraindications

None known.

Pre-use/Handling/Inspection

- Check the packaging for damage which can compromise sterility.
- Dispose of the sterile single-use vaginal speculums in accordance with all applicable federal, state and local medical and hazardous waste practices.
- Do not reuse or re-sterilize single use devices

Warnings

- Do not Reuse Single-Use Devices.
- Do not re-sterilize sterile packaged devices
- Do not use these instruments for purposes other than those for which they are intended.
- Do not allow conductive instruments to contact active cautery devices, or use in current paths, as patient or user burns may occur.
- Do not bend, pry, or use excessive force; breakage or failure of the instrument could occur resulting in possible harm to the patient or user.
- Use extreme care during handling of delicate or sharp instruments as injury or damage could occur.
- Use appropriate personal protection equipment when handling contaminated instruments.

Troubleshooting And Repairs

North EOS Industries Inc. disclaims all responsibility in cases of improper user handling or unauthorized repairs.

Returns

Contact Phoenix Surgical Supplies to obtain a Return Goods Authorization number (RGA#) prior to shipping the product back to North EOS Industries Inc. The RGA# should be prominently displayed on the outside of the return packaging and included on all paperwork enclosed with the return. All product returned should be decontaminated and cleaned, and all products should be safely packed in protective wrapping. Please have the original invoice number or purchase order number available to assist in verifying warranty information.

Customer Service Information

For further information regarding the use of this product or to report any problems, please contact Phoenix Surgical Supplies on 08 6311 2885

Limited Warranty

- A. This LIMITED WARRANTY provides assurance for the customer who purchases a North EOS Industries Inc Product (hereinafter the "Product") that should the Product fail to function to North EOS Industries Inc' published specifications during the term of this LIMITED WARRANTY (one year from the date of shipment for new Product), North EOS Industries Inc will either replace, repair, or issue a credit (adjusted to reflect the age of the Product) for the Product or any portion thereof. This LIMITED WARRANTY is extended only to the buyer purchasing the Product directly from North EOS Industries Inc or from its affiliate or its authorized distributor or representative.
- B. To qualify for this LIMITED WARRANTY, the following conditions must be met:
- (1). The Product must be used on or before its "Use By" or "Use Before" date, if applicable.
 - (2). The Product must be used in accordance with its labelling and may not be altered or subjected to misuse, abuse, accident or improper handling.
 - (3). North EOS Industries Inc must be notified in writing within thirty (30) days following discovery of a defect.
 - (4). The Product must be returned to North EOS Industries Inc within thirty (30) days of North EOS Industries Inc receiving notice as provided for in (3) above.
 - (5). Upon examination of the Product by Phoenix Surgical, North EOS Industries Inc shall have determined that:
 - (i) the Product was not repaired or altered by anyone other than North EOS Industries Inc or its authorized representative,
 - (ii) the Product was not operated under conditions other than normal use, and
 - (iii) the prescribed periodic maintenance and services have been performed on the Product.
- C. This LIMITED WARRANTY is limited to its express terms. THIS LIMITED WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED WHETHER STATUTORY OR OTHERWISE, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. In no event shall North EOS Industries Inc be liable for any consequential, incidental, prospective or other similar damage resulting from a defect, failure, or malfunction of the Product, whether a claim for such damage is based upon the warranty, contract, negligence or otherwise.
- D. The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. Users may benefit from statutory warranty rights under legislation governing the sale of consumer goods. If any part or term of this LIMITED WARRANTY is held by any court of competent jurisdiction to be illegal, unenforceable, or in conflict with applicable law, the validity of the remaining portion of the LIMITED WARRANTY shall not be affected, and all rights and obligations shall be construed and enforced as if this LIMITED WARRANTY did not contain the particular part or term held to be invalid.

The information contained in this document is accurate at time of publication. North EOS Industries Inc reserves the right to make changes to the product described in this manual without notice and without incorporating those changes to products already sold.

THANK YOU FOR YOUR ONGOING SUPPORT OF OUR AUSTRALIAN BUSINESS